

Certificate of Analysis

Product name: Alsevers Solution

Product code(s): 04-181 – 04-184

Batch Numbers: 32160201

Date Manufactured: 03 Feb 2016

Expiry Date: 02 Feb 2017

Appearance: Clear Liquid

Quality Control:

	Acceptable Ranges	Batch result
pH	6.80-7.30	6.81
Conductivity(μs/cm)	11500-13000	11540
Osmolarity(mOsm/kg)	335-360	335

Preservatives used: Chloramphenicol and Neomycin Sulphate

Sterility: Terminally filtered with 0.2 μ m filter in a grade C clean room and tested for sterility.

Storage: 2-8°C

Country of Origin: United Kingdom

Health and Safety:

Precautionary Notes

Chloramphenicol and Neomycin Sulphate are 'Toxic' and 'Harmful' substances retrospectively. Although present in this reagent in very small quantities, appropriate health and safety interventions are recommended. Exposure of preserved reagents directly to the skin and eyes should be avoided. Appropriate Personal Protective Equipment (PPE) should be worn at all times whilst carrying out tests. Accidental exposure to the skin should be treated by washing with mild detergent and water for 15-20 minutes. Eyes should be washed using water or normal saline for the same duration.

Signed for Release:

Date: 04 Feb 2016

