

## Certificate of Analysis

**Product name: Amnicryo Solution**

**Product code(s): 04-326**

**Batch Numbers: 72160504**

**Date Manufactured: 25 May 2016**

**Expiry Date: 24 May 2017**

**Appearance: Red Liquid**

**Quality Control:**

	<b>Acceptable Ranges</b>	<b>Batch result</b>
<b>pH</b>	7.00-7.50	7.20
<b>Conductivity(<math>\mu</math>s/cm)</b>	1250-1600	1438
<b>Osmolarity(mOsm/kg)</b>	1680-1850	1744

**Preservatives used:** None

**Sterility:** Terminally filtered with 0.2 $\mu$ m filter in a grade B clean room and tested for sterility.

**Storage:** Room temperature

**Country of Origin:** United Kingdom

**Health and Safety:**

*Precautionary Notes*

There are no specific Health and Safety problems associated with the use of Amnicryo Solution.

**Signed for Release:**

**Date: 05 Jun 2016**

