

Certificate of Analysis

Product name: Amnicryo Solution

Product code(s): 04-326

Batch Numbers: 72160505

Date Manufactured: 25 May 2016

Expiry Date: 24 May 2017

Appearance: Red Liquid

Quality Control:

	Acceptable Ranges	Batch result
pH	7.00-7.50	7.20
Conductivity(μs/cm)	1250-1600	1438
Osmolarity(mOsm/kg)	1680-1850	1744

Preservatives used: None

Sterility: Terminally filtered with 0.2 μ m filter in a grade B clean room and tested for sterility.

Storage: Room temperature

Country of Origin: United Kingdom

Health and Safety:

Precautionary Notes

There are no specific Health and Safety problems associated with the use of Amnicryo Solution.

Signed for Release:

Date: 05 Jun 2016

