

Certificate of Analysis

Product name: Heart Valve Antibiotic Solution (HVAS)

Product code(s): 04-341
Batch Numbers: 20161001
Date Manufactured: 10 Oct 2016
Expiry Date: 09 Oct 2017
Appearance: Red Liquid

Quality Control:

	Acceptable Ranges	Batch result
pH	7.20 – 7.40	7.11
Conductivity(µs/cm)	13000 - 15000	13150
Osmolarity(mOsm/kg)	280-300	299

Preservatives used: None

Sterility: Terminally filtered with 0.2µm filter in a grade B clean room and tested for sterility.

Storage: Store at below -15°C

Country of Origin: United Kingdom

Health and Safety:

Precautionary Notes

Gentamicin sulphate is classified as 'Harmful'. While this antibiotic is dissolved in the reagent, appropriate health and safety interventions are recommended. Appropriate Personal Protective Equipment should be worn at all times whilst carrying out tests. Accidental exposure to the skin should be treated by washing with mild detergent and water for 15-20 minutes. Eyes should be washed using water or normal saline for the same duration.

Signed for Release:

Date: 21 Oct 2016

