

Certificate of Analysis

Product name: Fixol
Fixative reagent for the FCD kit

Product code(s): 05-051

Batch Numbers: 51151112

Date Manufactured: 26/11/2015

Expiry Date: 25/11/2017

Appearance: Clear Fluid

Quality Control:

	Acceptable Ranges	Batch result
pH	8.50-9.50	9.07
Conductivity(μs/cm)	1750-2150	1800
Osmolarity(mOsm/kg)	N/A	N/A

Preservatives used: Alcohol Based

Sterility: Not required

Storage: Room temperature

Country of Origin: United Kingdom

Health and Safety:

Precautionary Notes

Methanol is a very toxic substance if swallowed. It can be fatal if swallowed. It is also toxic by inhalation. It is an irritant to the eyes.

Methanol is highly flammable.

It is strongly recommended that the user holds the appropriate COSHH or MSDS information on this product.

Signed for Quality Assurance:

Date: 03 Dec 2015

