

Certificate of Analysis

Product name: Fetal Cell Detection Kit (FCD)

Product code(s): 05-050 – 05-052

Batch Numbers: 05160810 A/B

Date Manufactured: 17 Aug 2016

Expiry Date: 16 Aug 2018

| | | |
|--------------------|----------------------|-------------------------|
| Appearance: | Fixative Reagent | Clear Fluid |
| | Eluting Reagent | Black/Deep Purple Fluid |
| | Counterstain Reagent | Bright Red Fluid |

Quality Control:

| | | | |
|------------|----------------------|-----------|-------------|
| pH: | Eluting Reagent | 1.80-2.10 | 1.83 |
| | Counterstain Reagent | N/A | |

| | | |
|----------------------|----------------------|-----|
| Conductivity: | Eluting Reagent | N/A |
| | Counterstain Reagent | N/A |

| | | |
|----------------------------|----------------------|-------------------|
| Preservatives used: | Eluting Reagent | Alcohol Based |
| | Counterstain Reagent | 0.05 Sodium azide |

Storage: Room temperature

Country of Origin: United Kingdom

Health and Safety:

Precautionary Notes

Methanol is a very toxic substance and can be fatal if swallowed. It is also toxic by inhalation, an irritant to the eyes and is highly flammable. Sodium azide is poisonous and is an irritant. It is fatal if swallowed and may cause cardiac disturbances. Although present in these reagents in small quantities, appropriate health and safety interventions are recommended.

It is strongly recommended that the user holds the appropriate COSHH or MSDS information these products. Appropriate Personal Protective Equipment should be worn at all times whilst carrying out tests. Accidental exposure to the skin should be treated by washing with mild detergent and water for 15-20 minutes. Eyes should be washed using water or normal saline for the same time.

Seek medical advice if any of these reagents are swallowed.

Signed for Quality Assurance:

Date: 26 Aug 2016

