

## Certificate of Analysis

**Product name:** Checkzyme

**Product code(s):** 04-037, 04-038 and 04-151

**Batch Numbers:** 15160803

**Date Manufactured:** 19 Aug 2016

**Expiry Date:** 18 Aug 2018

**Appearance:** Clear- Yellow coloured liquid

**Quality Control:**

	Acceptable Ranges	This batch result
pH	6.9-7.1	7.02
Conductivity( $\mu$ s/cm)	N/A	N/A
Osmolarity(mOsm/kg)	285-295	291

**Preservatives used:** Sodium azide (0.09% [w/v])

**Storage:** 2-8°C

**Country of Origin:** United Kingdom

### Health and Safety:

#### *Precautionary Notes*

Care should be taken if glass vials are present in this kit. In the case of breakage resulting in a sharps injury, the appropriate first aid measures should be taken.

Sodium azide is poisonous and is an irritant. It can be fatal if swallowed and may cause cardiac disturbances. Although present in these reagents in small quantities, appropriate health and safety interventions are recommended. Exposure of preserved reagents directly to the skin and eyes should be avoided. Appropriate Personal Protective Equipment should be worn at all times whilst carrying out tests. Accidental exposure to the skin should be treated by washing with mild detergent and water for 15-20 minutes. Eyes should be washed using water or normal saline for the same time.

**Signed for Release:**

**Date:** 21 Oct 2016

