

Certificate of Analysis

Product name: De-Epidermalising Reagent

Product code(s): 04-332

Batch Numbers: 75160403

Date Manufactured: 21 Apr 16

Expiry Date: 20 Apr 17

Appearance: Clear Liquid

Quality Control:

	Acceptable Ranges	Batch result
pH	N/A	N/A
Conductivity(μs/cm)	N/A	N/A
Osmolarity(mOsm/kg)	1820-1900	1841

Meets the requirements set out by Tissue Services, Liverpool.

Preservatives used: None

Sterility: Terminally filtered with 0.2 μ m filter in a grade B clean room and tested for sterility.

Storage: 2-8°C

Country of Origin: United Kingdom

Health and Safety:

Precautionary Notes

There are no Health and Safety adverse issues associated with this product.

Signed for Quality Assurance:

Date: 04 May 2016

