

## Certificate of Analysis

**Product name: Fetal Cell Detection Kit (FCD)**

**Product code(s): 05-050 – 05-052**

**Batch Numbers: 051612017 A/B**

**Date Manufactured: 02 Dec 2016**

**Expiry Date: 01 Dec 2018**

<b>Appearance:</b>	Fixative Reagent	Clear Fluid
	Eluting Reagent	Black/Deep Purple Fluid
	Counterstain Reagent	Bright Red Fluid

**Quality Control:**

<b>pH:</b>	Eluting Reagent	1.80-2.10	<b>1.91</b>
	Counterstain Reagent	N/A	

<b>Conductivity:</b>	Eluting Reagent	N/A
	Counterstain Reagent	N/A

<b>Preservatives used:</b>	Eluting Reagent	Alcohol Based
	Counterstain Reagent	0.05 Sodium azide

**Storage:** Room temperature

**Country of Origin:** United Kingdom

**Health and Safety:**

*Precautionary Notes*

Methanol is a very toxic substance and can be fatal if swallowed. It is also toxic by inhalation, an irritant to the eyes and is highly flammable. Sodium azide is poisonous and is an irritant. It is fatal if swallowed and may cause cardiac disturbances. Although present in these reagents in small quantities, appropriate health and safety interventions are recommended.

It is strongly recommended that the user holds the appropriate COSHH or MSDS information these products. Appropriate Personal Protective Equipment should be worn at all times whilst carrying out tests. Accidental exposure to the skin should be treated by washing with mild detergent and water for 15-20 minutes. Eyes should be washed using water or normal saline for the same time.

Seek medical advice if any of these reagents are swallowed.

**Signed for Quality Assurance:**

**Date: 13 Dec 2016**

