

## Certificate of Analysis

**Product name:** Filtered Reagent Grade Water

**Product code(s):** 04-036

**Batch Numbers:** 18170103

**Date Manufactured:** 30 Jan 2017

**Expiry Date:** 29 Jan 2018

**Appearance:** Clear liquid

**Quality Control:**

	Acceptable Ranges	Batch result
<b>pH</b>	N/A	N/A
<b>Conductivity(<math>\mu</math>s/cm)</b>	<10	9.40
<b>Osmolarity(mOsm/kg)</b>	N/A	N/A

**Preservatives used:** None

**Sterility:** Terminally filtered with 0.2 $\mu$ m filter in a grade B clean room and tested for sterility.

**Storage:** Store at room temperature

**Country of Origin:** United Kingdom

**Health and Safety:**

*Precautionary Notes*

There are no health and safety issues associated with this product.

**Signed for Quality Assurance:**

**Date:** 10 Feb 2017

