

Certificate of Analysis

Product name: Fixol
Fixative reagent for the FCD kit

Product code(s): 05-051

Batch Numbers: 51170124

Date Manufactured: 25 Jan 2016

Expiry Date: 24 Jan 2019

Appearance: Clear Fluid

Quality Control:

	Acceptable Ranges	Batch result
pH	8.50-9.50	9.04
Conductivity(μs/cm)	1750-2150	2100
Osmolarity(mOsm/kg)	N/A	N/A

Preservatives used: Alcohol Based

Sterility: Not required

Storage: Room temperature

Country of Origin: United Kingdom

Health and Safety:

Precautionary Notes

Methanol is a very toxic substance if swallowed. It can be fatal if swallowed. It is also toxic by inhalation. It is an irritant to the eyes.

Methanol is highly flammable.

It is strongly recommended that the user holds the appropriate COSHH or MSDS information on this product.

Signed for Quality Assurance:

Date: 01 Mar 2017

