

Certificate of Analysis

Product name: LISS 2 (LISS Additive Reagent)

Product code(s): 04-070 and 04-076 – 04-078 and 04-180

Batch Numbers: 34161207

Date Manufactured: 14 Dec 2016

Expiry Date: 13 Dec 2018

Appearance: Clear Liquid

Quality Control:

	Acceptable Ranges	Batch result
pH	6.70-7.10	7.01
Conductivity(μs/cm)	3000-3600	3353
Osmolarity(mOsm/kg)	290-310	310

Preservatives used: Sodium azide (0.09% [w/v])

Sterility: Terminally filtered with 0.2 μ m filter in a grade B clean room and tested for sterility.

Storage: 2-8°C

Country of Origin: United Kingdom

Health and Safety:

Precautionary Notes

Sodium azide is poisonous and is an irritant. It is fatal if swallowed and may cause cardiac disturbances. Although present in these reagents in small quantities, appropriate health and safety interventions are recommended. Exposure of preserved reagents directly to the skin and eyes should be avoided. Appropriate Personal Protective Equipment should be worn at all times whilst carrying out tests. Accidental exposure to the skin should be treated by washing with mild detergent and water for 15-20 minutes. Eyes should be washed using water or normal saline for the same duration.

Signed for Quality Assurance:

Date:

03 Jan 2017

