

Certificate of Analysis

**Product name: POLYAGGLUTINATION
KIT**

Product code(s): 03-061

Batch Numbers: 79170408

Date Manufactured: 04 Apr 2017

Expiry Date: 18 Nov 2017

Appearance: Clear – pale brown liquid

Kit made up of four lectin extracts:

- 1 Arachis hypogea - **14160903**
- 2 Glycine soja - **15151103**
- 3 Salvia Sclarea - **79160803**
- 4 Dolichos biflorus - **29161010**

Quality Control:

No specific guidelines are laid down for a polyagglutination kit but the kit is expected to show a pattern of agglutination with affected human red cells which will classify them as being T, Tn, Tk or Cad. It may be possible to further determine Th and Tx affected red cells.

Preservatives used: Sodium azide (0.1% [w/v])

Country of Origin: United Kingdom

Health and Safety:

Precautionary Notes

Sodium azide is poisonous and is an irritant. It is fatal if swallowed and may cause cardiac disturbances. Although present in these reagents in very small quantities, appropriate health and safety interventions are recommended. Appropriate Personal Protective Equipment should be worn at all times whilst carrying out tests. Accidental exposure to the skin should be treated by washing with mild detergent and water for 15-20 minutes. Eyes should be washed using water or normal saline for the same duration.

Signed for Release:

Date: 04 Apr 2017

