

Certificate of Analysis

Product name: Reagent Grade Water

Product code(s): 04-035

Batch Numbers: 17161153

Date Manufactured: 10 Nov 16

Expiry Date: 09 Nov 17

Appearance: Clear

Quality Control:

	Acceptable Ranges	Batch result
pH	N/A	N/A
Conductivity(μ s/cm)	<10	8.35
Osmolarity(mOsm/kg)	N/A	N/A

Preservatives used: None

Sterility: No chemicals used in this preparation. Water is run through the reverse osmosis unit and collected without it being stored. It is collected straight into the appropriate bags or bottles with as little exposure to the atmosphere as possible.

Storage: Store at room temperature

Country of Origin: United Kingdom

Health and Safety:
Precautionary Notes

Signed for Release:

Date: 10 Nov 16

