

## Certificate of Analysis

**Product name: Fetal Cell Detection Kit (FCD)**

**Product code(s): 05-050 – 05-052**

**Batch Numbers: 05170421 A/B**

**Date Manufactured: Eluting: 19 Apr 2017 / Counterstain: 02 May 2017**

**Kit Expiry Date: 18 Apr 2019**

**Appearance:** Eluting Reagent                      Black/Deep Purple Fluid  
Counterstain Reagent                      Bright Red Fluid

Quality Control:	pH	Range	This Batch
	Eluting Reagent:		1.80-2.10
Counterstain Reagent:		N/A	N/A

The performance of the above components has been assessed by Kleihauer testing of negative and positive control samples. All tests met the performance specifications for this product and passed the assessment.

[Demonstration of fetal hemoglobin in erythrocytes of a blood smear].

KLEIHAUER E, BRAUN H, BETKE K  
Klin Wochenschr 1957 (35):637-638

**Preservatives used:** Eluting Reagent                      Alcohol Based  
Counterstain Reagent                      0.10% w/v Sodium azide

**Storage:** Room temperature

**Country of Origin:** United Kingdom

**Health and Safety:** *Precautionary Notes*

Find the Safety Data sheet on our website at [www.sourcebioscience.com](http://www.sourcebioscience.com)

Appropriate Personal Protective Equipment should be worn at all times whilst carrying out tests.  
**Seek medical advice if any of these reagents are swallowed.**

**Signed for Quality Assurance:**

**Date: 16 May 2017**

