

## Certificate of Analysis

**Product name: Fixol**  
Fixative reagent for the FCD kit

**Product code(s): 05-051**

**Batch Numbers: 51170427**

**Date Manufactured: 02 May 2017**

**Expiry Date: 01 May 2019**

**Appearance: Clear Fluid**

Quality Control:	Acceptable Ranges	Batch result
pH	8.50-9.50	8.97
Conductivity( $\mu\text{s}/\text{cm}$ )	1750-2150	2150

The performance of the above components has been assessed by Kleihauer testing of negative and positive control samples. All tests met the performance specifications for this product and passed the assessment.

[Demonstration of fetal hemoglobin in erythrocytes of a blood smear].  
KLEIHAUER E, BRAUN H, BETKE K  
Klin Wochenschr 1957 (35):637-638

**Preservatives used:** Alcohol Based

**Sterility:** Not required

**Storage:** Room temperature

**Country of Origin:** United Kingdom

**Health and Safety:** *Precautionary Notes*

Find the Safety Data sheet on our website at [www.sourcebioscience.com](http://www.sourcebioscience.com)

Appropriate Personal Protective Equipment should be worn at all times whilst carrying out tests.  
**Seek medical advice if any of these reagents are swallowed.**

**Signed for Quality Assurance:**

**Date: 16 May 2017**

