

Certificate of Analysis

Product name: Reagent Grade Water

Product code(s): 04-035

Batch Numbers: 17170567

Date Manufactured: 25 May 17

Expiry Date: 24 May 18

Appearance: Clear

Quality Control:

	Acceptable Ranges	Batch result
pH	N/A	N/A
Conductivity(μ s/cm)	<10	5.90
Osmolarity(mOsm/kg)	N/A	N/A

Preservatives used: None

Sterility: No chemicals used in this preparation. Water is run through the reverse osmosis unit and collected without it being stored. It is collected straight into the appropriate bags with as little exposure to the atmosphere as possible.

Storage: Store at room temperature

Country of Origin: United Kingdom

Health and Safety:

No health and safety implications from this product;

Download our safety data sheet from www.sourcebioscience.com

Signed for Release:

Date: 25 May 2017

