

CERTIFICATE OF ANALYSIS

Certificate of Analysis

Product name: Modified Alsever's Solution

Product code(s): 04-182

Batch Number(s): 32170713

Date Manufactured: 24 Jul 2017

Expiry Date: 23 Jul 2018

Appearance: Clear liquid

Quality Control:

	Specification	This batch result
pH	6.80 - 7.30	7.12
Conductivity($\mu\text{s}/\text{cm}$)	11500 - 13000	12480
Osmolality(mOsm/kg)	335 - 360	351

Preservatives used: Chloramphenicol and Neomycin sulphate

Sterility: Terminally filtered through a 0.2 μm filter in a grade B clean room and sterility tested.

Storage: 2 - 8 °C

Country of Origin: United Kingdom

Health and Safety:

Precautionary Notes

Appropriate Personal Protective Equipment (PPE) should be worn at all times whilst carrying out tests. Please download the Safety Data Sheet at <http://sourcebioscience.com>

Signed for Quality Control:



Date:

02 Aug 17

Signed for Certificate Check:



Date:

02 Aug 17