

## CERTIFICATE OF ANALYSIS

# Certificate of Analysis

**Product name:** POLYAGGLUTINATION KIT  
**Product code(s):** 03-061  
**Batch Number(s):** 79170810  
**Date Manufactured:** 23 Aug 17  
**Expiry Date:** 16 Aug 19  
**Appearance:** Clear – pale brown liquid  
Kit made up of four lectin extracts: 1. Arachis hypogea - 14170806, 2. Glycine soja - 15170806  
3. Salvia Sclarea - 79170809, 4. Dolichos biflorus - 29170818

### Quality Control:

	Specification	This batch result
pH	N/A	N/A
Conductivity( $\mu\text{s}/\text{cm}$ )	N/A	N/A
Osmolality(mOsm/kg)	N/A	N/A

No specific guidelines are laid down for a polyagglutination kit but the kit is expected to show a pattern of agglutination with affected human red cells which will classify them as being T, Tn, Tk or Cad. It may be possible to further determine Th and Tx affected red cells

**Preservatives used:** Sodium azide (0.1% [w/v])

**Sterility:** N/A

**Storage:** 2 - 8 °C

**Country of Origin:** United Kingdom

### Health and Safety:

#### Precautionary Notes

Please download the Safety Data Sheet at <http://sourcebioscience.com>

Signed for Quality Control: 

Date: 26 Aug 17

Signed for Certificate Check: 

Date: 26 Aug 17