

CERTIFICATE OF ANALYSIS

Certificate of Analysis

Product name: Fixol (Fixative reagent for Fetal Cell Detection Kit)
Product code(s): 05-051
Batch Number(s): 51170629
Date Manufactured: 06 Jun 2017
Expiry Date: 05 Jun 2019
Appearance: Colourless transparent liquid

Quality Control:

	Specification	This batch result
pH	8.50 - 9.50	8.87
Conductivity(μs/cm)	1750 - 2150	2140
Osmolality(mOsm/kg)	N/A	N/A

The performance of the above components has been assessed by Kleihauer testing of negative and positive control samples. All tests met the performance specifications for this product and passed the assessment. [Demonstration of fetal hemoglobin in erythrocytes of a blood smear]. KLEIHAUER E, BRAUN H, BETKE K Klin Wochenschr 1957 (35):637-638

Preservatives used: Sodium azide (0.024% w/v), Alcohol based

Sterility: N/A
Storage: Room Temperature

Country of Origin: United Kingdom

Health and Safety:

Precautionary Notes

Please download the Safety Data Sheet at <http://sourcebioscience.com>

Signed for Quality Control:



Date:

21 Jul 17

Signed for Certificate Check:



Date:

21 Jul 17