

CERTIFICATE OF ANALYSIS

Certificate of Analysis

Product name: Checkzyme (Glycine soja extract)
Product code(s): 04-037, 04-037/100, 04-038 and 04-151
Batch Number(s): 15170806
Date Manufactured: 21 Aug 17
Expiry Date: 16 Aug 19
Appearance: Clear – Yellow Coloured liquid

Quality Control:

	Specification	This batch result
pH	6.9-7.1	7.06
Conductivity($\mu\text{s}/\text{cm}$)	N/A	N/A
Osmolality(mOsm/kg)	285-1000	294

Preservatives used: Sodium azide (0.1% [w/v])

Sterility: N/A

Storage: 2 - 8 °C

Country of Origin: United Kingdom

Health and Safety:

Precautionary Notes

Sodium azide is poisonous and is an irritant. It can be fatal if swallowed and may cause cardiac disturbances. Although present in these reagents in small quantities, appropriate health and safety interventions are recommended. Appropriate Personal Protective Equipment should be worn at all times whilst carrying out tests. Accidental exposure to the skin should be treated by washing with mild detergent and water for 15-20 minutes. Eyes should be washed using water or normal saline for the same duration.

Signed for Quality Control: 

Date: 25 Aug 17

Signed for Certificate Check: 

Date: 25 Aug 17