

CERTIFICATE OF ANALYSIS

Certificate of Analysis

Product name: PBSS (Phosphate Buffered Serology Saline)

Product code(s): 04-004

Batch Number(s): 01170913

Date Manufactured: 02 Sep 2017

Expiry Date: 01 Aug 2019

Appearance: Clear Liquid

Quality Control:

	Specification	This batch result
pH	6.95-7.05	7.04
Conductivity($\mu\text{s}/\text{cm}$)	N/A	N/A
Osmolality(mOsm/kg)	285-295	288

Meets the requirements set out in 'Guidelines for the Blood Transfusion Services in the United Kingdom' 7th Edition; as set out in chapter 12, 12.2.4, Miscellaneous reagents.

Preservatives used: Sodium azide (0.01% [w/v])

Sterility: Product is filtered through a 0.2 μm filter into gamma irradiated bags

Storage: Room temperature

Country of Origin: United Kingdom

Health and Safety:

Precautionary Notes

Please download the Safety Data Sheet at <http://sourcebioscience.com>

Signed for Quality Control: 

Date: 06 Sep 17

Signed for Certificate Check: 

Date: 04 Sep 17