

## CERTIFICATE OF ANALYSIS

# Certificate of Analysis

**Product name:** TBIS (Tris Buffered Isotonic Saline)

**Product code(s):** 04-142

**Batch Number(s):** 37170806

**Date Manufactured:** 30 Aug 17

**Expiry Date:** 29 Aug 18

**Appearance:** Clear liquid

**Quality Control:**

	Specification	This batch result
pH	7.30 - 7.60	7.48
Conductivity( $\mu\text{s}/\text{cm}$ )	14000 - 16500	16490
Osmolality(mOsm/kg)	285 - 300	292

**Preservatives used:** None.

**Sterility:** Terminally filtered with a 0.2 $\mu\text{m}$  filter in a grade B clean room and tested for sterility.

**Storage:** Room Temperature

**Country of Origin:** United Kingdom

**Health and Safety:**

**Precautionary Notes**

Although there are no specific concerns please use appropriate PPE when handling this reagent. Please download the Safety Data Sheet at <http://sourcebioscience.com>

**Signed for Quality Control:** 

**Date:** 08 Sep 17

**Signed for Certificate Check:** 

**Date:** 08 Sep 17