

## CERTIFICATE OF ANALYSIS

# Certificate of Analysis

**Product name:** PBSS (Phosphate Buffered Serology Saline)

**Product code(s):** 04-004

**Batch Number(s):** 01170929

**Date Manufactured:** 12 Sep 2017

**Expiry Date:** 11 Sep 2019

**Appearance:** Clear Liquid

**Quality Control:**

	Specification	This batch result
pH	6.95-7.05	7.04
Conductivity( $\mu\text{s}/\text{cm}$ )	N/A	N/A
Osmolality(mOsm/kg)	285-295	294

Meets the requirements set out in 'Guidelines for the Blood Transfusion Services in the United Kingdom' 7th Edition; as set out in chapter 12, 12.2.4, Miscellaneous reagents.

**Preservatives used:** Sodium azide (0.01% [w/v])

**Sterility:** Product is filtered through a 0.2 $\mu\text{m}$  filter into gamma irradiated bags

**Storage:** Room temperature

**Country of Origin:** United Kingdom

**Health and Safety:**

**Precautionary Notes**

Please download the Safety Data Sheet at <http://sourcebioscience.com>

Signed for Quality Control: 

Date: 14 Sep 17

Signed for Certificate Check: 

Date: 14 Sep 17