

## CERTIFICATE OF ANALYSIS

# Certificate of Analysis

**Product name:** Cryoprotectant Reagent

**Product code(s):** 04-213PH

**Batch Number(s):** 61171107

**Date Manufactured:** 14 Nov 17

**Expiry Date:** 13 Nov 19

**Appearance:** Clear Liquid

**Quality Control:**

	Specification	This batch result
pH	5.8-6.7	5.81
Conductivity( $\mu\text{s}/\text{cm}$ )	4200-4850	4486
Osmolality(mOsm/kg)	880-970	920

**Preservatives used:** None

**Sterility:** Terminally filtered with 0.2 $\mu\text{m}$  filter in a grade B clean room and tested for sterility.

**Storage:** Ambient

**Country of Origin:** United Kingdom

**Health and Safety:**

**Precautionary Notes**

Please download the Safety Data Sheet at <http://sourcebioscience.com>

**Signed for Quality Control:**



**Date:** 22 NOV 17

**Signed for Certificate Check:**



**Date:** 22 NOV 17