

## CERTIFICATE OF ANALYSIS

# Certificate of Analysis

**Product name:** PBSS (Phosphate Buffered Serology Saline)

**Product code(s):** 04-004

**Batch Number(s):** 01171190

**Date Manufactured:** 06 Nov 17

**Expiry Date:** 05 Nov 19

**Appearance:** Clear Liquid

**Quality Control:**

|   | Specification | This batch result |
|---|---------------|-------------------|
| pH                                      | 6.95-7.05     | 7.05              |
| Conductivity( $\mu\text{s}/\text{cm}$ ) | N/A           | N/A               |
| Osmolality(mOsm/kg)                     | 285-295       | 294               |

Meets the requirements set out in 'Guidelines for the Blood Transfusion Services in the United Kingdom' 7th Edition; as set out in chapter 12, 12.2.4, Miscellaneous reagents.

**Preservatives used:** Sodium azide (0.01% [w/v])

**Sterility:** Product is filtered through a 0.2 $\mu\text{m}$  filter into gamma irradiated bags

**Storage:** Room temperature

**Country of Origin:** United Kingdom

**Health and Safety:**

**Precautionary Notes**

Please download the Safety Data Sheet at <http://sourcebioscience.com>

**Signed for Quality Control:**



**Date:** 29 NOV 17

**Signed for Certificate Check:**



**Date:** 29 NOV 17