

CERTIFICATE OF ANALYSIS

Certificate of Analysis

Product name: TBIS (Tris Buffered Isotonic Saline)

Product code(s): 04-142

Batch Number(s): 37171208

Date Manufactured: 18 Dec 17

Expiry Date: 17 Dec 18

Appearance: Clear liquid

Quality Control:

	Specification	This batch result
pH	7.30 - 7.60	7.56
Conductivity($\mu\text{s}/\text{cm}$)	14000 - 16500	15630
Osmolality(mOsm/kg)	285 - 300	286

Preservatives used: None.

Sterility: Terminally filtered with a 0.2 μm filter in a grade B clean room and tested for sterility.

Storage: Room Temperature

Country of Origin: United Kingdom

Health and Safety:

Precautionary Notes


Although there are no specific concerns please use appropriate PPE when handling this reagent. Please download the Safety Data Sheet at <http://sourcebioscience.com>

Signed for Quality Control:



Date: 06 Jan 18

Signed for Certificate Check:



Date: 05 Jan 18