

CERTIFICATE OF ANALYSIS

Certificate of Analysis

Product name: PBSS (Phosphate Buffered Serology Saline)

Product code(s): 04-004
Batch Number(s): 01180132
Date Manufactured: 10 Jan 2018
Expiry Date: 09 Jan 2020
Appearance: Clear Liquid

Quality Control:

| | Specification | This batch result |
|---------------------------|---------------|-------------------|
| pH | 6.95-7.05 | 7.04 |
| Conductivity(μ s/cm) | N/A | N/A |
| Osmolality(mOsm/kg) | 285-295 | 295 |

e.g. Meets the requirements set out in 'Guidelines for the Blood Transfusion Services in the United Kingdom' 7th Edition; as set out in chapter 12, 12.2.4, Miscellaneous reagents.

Preservatives used: Sodium azide (0.01% [w/v])

Sterility: Product is filtered through a 0.2 μ m filter into gamma irradiated bags

Storage: Room temperature

Country of Origin: United Kingdom

Health and Safety:

Precautionary Notes

Please download the Safety Data Sheet at <http://sourcebioscience.com>

Signed for Quality Control: 

Date: 11 JAN 18.

Signed for Certificate Check: 

Date: 11 JAN 18