

CERTIFICATE OF ANALYSIS

Certificate of Analysis

Product name: Fixol (Fixative reagent for Fetal Cell Detection Kit)
Product code(s): 05-051
Batch Number(s): 51171135
Date Manufactured: 13 Dec 17
Expiry Date: 12 Dec 19
Appearance: Colourless transparent liquid

Quality Control:

	Specification	This batch result
pH	8.50-9.50	8.80
Conductivity($\mu\text{s}/\text{cm}$)	1750-2150	2070
Osmolality(mOsm/kg)	N/A	N/A

The performance of the above components has been assessed by Kleihauer testing of positive control samples. All tests met the performance specifications for this product and passed the assessment. [Demonstration of fetal hemoglobin in erythrocytes of a blood smear]. KLEIHAUER E, BRAUN H, BETKE K Klin Wochenschr 1957 (35):637-638

Preservatives used: Sodium azide (0.024% w/v), Alcohol based

Sterility: N/A

Storage: Room Temperature

Country of Origin: United Kingdom

Health and Safety:

Precautionary Notes

Please download the Safety Data Sheet at <http://sourcebioscience.com>

Signed for Quality Control:



Date: 22 Dec 17

Signed for Certificate Check:



Date: 22 Dec 17