

CERTIFICATE OF ANALYSIS

Certificate of Analysis

Product name: Fetal Cell Detection (FCD) Kit
Product code(s): 05-050
Batch Number(s): 05180130A/B
Date Manufactured: 05 Jan 18
Expiry Date: 03 Jan 20
Appearance: Eluting Reagent: Black/Brown Opaque Liquid
Counterstain Reagent: Pink/Red Translucent Liquid

Quality Control:

	Specification	This batch result
pH	Eluting: 1.80 - 2.10	2.01
Conductivity($\mu\text{s}/\text{cm}$)	N/A	N/A
Osmolality(mOsm/kg)	N/A	N/A

The performance of the above components has been assessed by Kleihauer testing of positive control samples. All tests met the performance specifications for this product and passed the assessment. [Demonstration of fetal hemoglobin in erythrocytes of a blood smear]. KLEIHAUER E, BRAUN H, BETKE K Klin Wochenschr 1957 (35):637-638

Preservatives used: Counterstain: Sodium azide (0.10% w/v), Eluting: Alcohol based

Sterility: N/A

Storage: Room temperature

Country of Origin: United Kingdom

Health and Safety:

Precautionary Notes

Please download the Safety Data Sheet at <http://sourcebioscience.com>

Signed for Quality Control:



Date: 17 Jan 18

Signed for Certificate Check:



Date: 17 Jan 18