

CERTIFICATE OF ANALYSIS

Certificate of Analysis

Product name: (Phosphate Buffered Serology Saline)

Product code(s): 04-004
Batch Number(s): 01180145
Date Manufactured: 19 Jan 2018
Expiry Date: 18 Jan 2020
Appearance: Clear Liquid

Quality Control:

	Specification	This batch result
pH	6.95-7.05	7.02
Conductivity(μ s/cm)	N/A	N/A
Osmolality(mOsm/kg)	285-295	288

e.g. Meets the requirements set out in 'Guidelines for the Blood Transfusion Services in the United Kingdom' 7th Edition; as set out in chapter 12, 12.2.4, Miscellaneous reagents.

Preservatives used: Sodium azide (0.01% [w/v])

Sterility: Product is filtered through a 0.2 μ m filter into gamma irradiated bags

Storage: Room temperature

Country of Origin: United Kingdom

Health and Safety:

Precautionary Notes

Please download the Safety Data Sheet at <http://sourcebioscience.com>

Signed for Quality Control: 

Date: 24 - JAN - 18

Signed for Certificate Check: 

Date: 29 JAN 18