

## CERTIFICATE OF ANALYSIS

# Certificate of Analysis

**Product name:** Reagent Grade Water

**Product code(s):** 04-035

**Batch Number(s):** 18180283

**Date Manufactured:** 08 Feb 2018

**Expiry Date:** 07 Feb 2019

**Appearance:** Clear Liquid

**Quality Control:**

	Specification	This batch result
pH	6.95-7.05	N/A
Conductivity( $\mu\text{s}/\text{cm}$ )	<10	4.87
Osmolality(mOsm/kg)	285-295	N/A

Meets the requirements set out in 'Guidelines for the Blood Transfusion Services in the United Kingdom' 7th Edition; as set out in chapter 12, 12.2.4, Miscellaneous reagents.

**Preservatives used:** N/A

**Sterility:** Product is filtered through a 0.2 $\mu\text{m}$  filter into gamma irradiated bags

**Storage:** Room temperature

**Country of Origin:** United Kingdom

**Health and Safety:**

**Precautionary Notes**

Please download the Safety Data Sheet at <http://sourcebioscience.com>

**Signed for Quality Control:** 

**Date:** 08 Feb 18

**Signed for Certificate Check:** 

**Date:** 08 Feb 18