

## CERTIFICATE OF ANALYSIS

# Certificate of Analysis

**Product name:** LISS 2 (LISS Additive Reagent)  
**Product code(s):** 04-070  
**Batch Number(s):** 34190114  
**Date Manufactured:** 16 Jan 19  
**Expiry Date:** 29 Mar 20  
**Appearance:** Clear Liquid

**Quality Control:**

	Specification	This batch result
pH	6.70-7.10	6.91
Conductivity( $\mu\text{s}/\text{cm}$ )	3000-3600	3400
Osmolality(mOsm/kg)	290-310	309

Meets the requirements set out in 'Guidelines for the Blood Transfusion Services in the United Kingdom' 7th Edition; as set out in chapter 12, 12.2.4, Miscellaneous reagents.

**Preservatives used:** Sodium azide (0.01% [w/v])

**Sterility:** Product is filtered through a 0.2 $\mu\text{m}$  and tested for micro sterility.

**Storage:** 2-8 °C

**Country of Origin:** United Kingdom

**Health and Safety:**

**Precautionary Notes**

Please download the Safety Data Sheet at <http://sourcebioscience.com>

**Signed for Quality Control:**



**Date:** 22 Feb 19

**Signed for Certificate Check:**



**Date:** 22 FEB 19