

CERTIFICATE OF ANALYSIS

Certificate of Analysis

Product name: Fixol (Fixative reagent for Fetal Cell Detection Kit) for guest medical

Product code(s): 05-051

Batch Number(s): 51190149C (Guest)

Date Manufactured: 28 Jan 19

Expiry Date: 23 Oct 20

Appearance: Clear transparent liquid

Quality Control:

	Specification	This batch result
pH	8.50 - 9.50	8.91
Conductivity($\mu\text{s}/\text{cm}$)	1750 - 2150	1968
Osmolality(mOsm/kg)	N/A	N/A

This product has been assessed and Passed using the Kleihaur testing of negative and positive control samples. (Demonstration of fetal hemoglobin in erythrocytes of a blood smear). KLEIHAUR E, BRAUN H, BETKE K Klin Wochenschr 1957 (35):637-638

Preservatives used: Methanol based product

Sterility: N/A

Storage: Ambient (Room Temperature)

Country of Origin: United Kingdom

Health and Safety:

Precautionary Notes

Please download the Safety Data Sheet at <http://sourcebioscience.com>

Signed for Quality Control:



Date: 26 Feb 19

Signed for Certificate Check:



Date: 26 Feb 19