

CERTIFICATE OF ANALYSIS

Certificate of Analysis

Product name: Low Ionic Strength Saline (LISS)

Product code(s): 04-123, 04-017

Batch Number(s): 03190331

Date Manufactured: 11 Mar 19

Expiry Date: 10 Mar 20

Appearance: Clear Liquid

Quality Control:

	Specification	This batch result
pH	6.50-6.90	6.80
Conductivity($\mu\text{s}/\text{cm}$)	3500-3800	3693
Osmolality(mOsm/kg)	285-295	290

Meets the requirements set out in 'Guidelines for the Blood Transfusion Services in the United Kingdom' 7th Edition; as set out in chapter 12, 12.2.4, Miscellaneous reagents.

Preservatives used: Sodium azide (0.01% [w/v])

Sterility: N/A

Storage: Room temperature

Country of Origin: United Kingdom

Health and Safety:

Precautionary Notes

Please download the Safety Data Sheet at <http://sourcebioscience.com>

Signed for Quality Control: 

Date: 19 MAR 19

Signed for Certificate Check: 

Date: 09 MAR 19