

CERTIFICATE OF ANALYSIS

Certificate of Analysis

Product name: Modified Alsevers Solution

Product code(s): 04-182

Batch Number(s): 32190734

Date Manufactured: 24 Jul 19

Expiry Date: 15 Jul 20

Appearance: Clear Liquid

Quality Control:

	Specification	This batch result
pH	6.80 - 7.30	7.19
Conductivity($\mu\text{s}/\text{cm}$)	11500 - 13000	11873
Osmolality(mOsm/kg)	N/A	N/A

Manufactured to SPN870/2.1 Dated 18/08/16

Preservatives used: Chloramphenicol and Neomycin sulphate

Sterility: Product is filtered through a 0.2 μm filter into gamma irradiated bags

Storage: 2-8 $^{\circ}\text{C}$

Country of Origin: United Kingdom

Health and Safety:

Precautionary Notes

Please download the Safety Data Sheet at <http://sourcebioscience.com>

Signed for Quality Control: 

Date: 01 Aug 19

Signed for Certificate Check: 

Date: 01 Aug 19