

CERTIFICATE OF ANALYSIS

Certificate of Analysis

Product name: Fixol (Fixative Reagent for Fetal Cell Detection Kit)

Product code(s): 05 - 051

Batch Number(s): 51190757 C

Date Manufactured: 23 Jul 19

Expiry Date: 03 Apr 21

Appearance: Clear Liquid

Quality Control:

	Specification	This batch result
pH	8.50 - 9.50	9.08
Conductivity($\mu\text{s}/\text{cm}$)	1750 - 2150	1895
Osmolality(mOsm/kg)	N/A	N/A

The performance of the above reagent has been assessed by Kleihauer testing of negative and positive control samples. All tests met the performance specifications for this product and passed the assessment. [Demonstration of fetal hemoglobin in erythrocytes of a blood smear]. KLEIHAUER E, BRAUN H, BETKE K Klin Wochenschr 1957 (35):637-638

Preservatives used: Methanol based product

Sterility: N/A

Storage: Room Temperature

Country of Origin: United Kingdom

Health and Safety:

Precautionary Notes

Please download the Safety Data Sheet at <http://sourcebioscience.com>

Signed for Quality Control:



Date: 14 AUG 19

Signed for Certificate Check:



Date: 14 Aug 19