

CERTIFICATE OF ANALYSIS

Certificate of Analysis

Product name: Cryoprotectant Reagent (CPR)

Product code(s): 04-213

Batch Number(s): 61190817

Date Manufactured: 21 Aug 19

Expiry Date: 17 Aug 20

Appearance: Clear Liquid

Quality Control:

	Specification	This batch result
pH	5.80-6.70	6.69
Conductivity(μ s/cm)	4200-4850	4819
Osmolality(mOsm/kg)	880-970	935

Preservatives used: N/A

Sterility: Terminally filtered with 0.2 μ m filter in a clean room and tested for sterility

Storage: Room temperature

Country of Origin: United Kingdom

Health and Safety:

Precautionary Notes

Please download the Safety Data Sheet at <http://sourcebioscience.com>

Signed for Quality Control:



Date: 05 Sep 19

Signed for Certificate Check:



Date: 05 Sep 19