

CERTIFICATE OF ANALYSIS

Certificate of Analysis

Product name: PBSS (Phosphate Buffered Serology Saline)

Product code(s): 04 - 004

Batch Number(s): 01190905

Date Manufactured: 09 SEP 2019

Expiry Date: 09 SEP 2021

Appearance: Clear Liquid

Quality Control:

	Specification	This batch result
pH	e.g. 6.95-7.05	6.97
Conductivity(μ s/cm)	e.g. N/A	N/a
Osmolality(mOsm/kg)	e.g. 285-295	292

Meets the requirements set out in 'Guidelines for the Blood Transfusion Services in the United Kingdom' 7th Edition; as set out in chapter 12, 12.2.4, Miscellaneous reagents.

Preservatives used: Sodium azide (0.01% [w/v])

Sterility: Product is filled in Gamma Irradiated bags

Storage: Room temperature

Country of Origin: United Kingdom

Health and Safety:

Precautionary Notes

Please download the Safety Data Sheet at <http://sourcebioscience.com>

Signed for Quality Control:



Date: 12 SEP 2019

Signed for Certificate Check:



Date: 12 sep 19