

CERTIFICATE OF ANALYSIS

Certificate of Analysis

Product name: Fetal Cell Detection (FCD) Kit
Product code(s): 05-050
Batch Number(s): 05190959A/B (Guest)
Date Manufactured: 04 Sep 19
Expiry Date: 24 Mar 21
Appearance: Eluting: Black/Brown opaque liquid
Counterstain: Pink/Red Translucent liquid

Quality Control:

	Specification	This batch result
pH	Eluting: 1.80 - 2.10	Eluting: 1.96
Conductivity($\mu\text{s}/\text{cm}$)	N/A	N/A
Osmolality(mOsm/kg)	N/A	N/A

This product has been assessed and Passed using the Kleihaur testing of negative and positive control samples. (Demonstration of fetal hemoglobin in erythrocytes of a blood smear). KLEIHAUR E, BRAUN H, BETKE K Klin Wochenschr 1957 (35):637-638

Preservatives used: Eluting: Alcohol based
Counterstain: Sodium Azide (0.10% w/v)

Sterility: N/A

Storage: Ambient (Room Temperature)

Country of Origin: United Kingdom

Health and Safety:

Precautionary Notes

Please download the Safety Data Sheet at <http://sourcebioscience.com>

Signed for Quality Control: 

Date: 15 Oct 19

Signed for Certificate Check: 

Date: 15 Oct 19