

CERTIFICATE OF ANALYSIS

Certificate of Analysis

Product name: TBIS (Tris Buffered Isotonic Saline)
Product code(s): 04-142
Batch Number(s): 37180308
Date Manufactured: 19 Mar 18
Expiry Date: 18 Mar 19
Appearance: Clear liquid

Quality Control:

	Specification	This batch result
pH	7.30 - 7.60	7.53
Conductivity($\mu\text{s}/\text{cm}$)	14000 - 16500	14856
Osmolality(mOsm/kg)	285 - 300	298

Preservatives used: None.

Sterility: Terminally filtered with a 0.2 μm filter in a grade B clean room and tested for sterility.

Storage: Room Temperature

Country of Origin: United Kingdom

Health and Safety:

Precautionary Notes

Although there are no specific concerns please use appropriate PPE when handling this reagent. Please download the Safety Data Sheet at <http://sourcebioscience.com>

Signed for Quality Control:  **Date:** 03 Apr 18

Signed for Certificate Check:  **Date:** 03 Apr 18