

## CERTIFICATE OF ANALYSIS

# Certificate of Analysis

**Product name:** PBSS (Phosphate Buffered Serology Saline)

**Product code(s):** 04-004

**Batch Number(s):** 01180334

**Date Manufactured:** 27 Mar 2018

**Expiry Date:** 26 Mar 2020

**Appearance:** Clear Liquid

**Quality Control:**

	Specification	This batch result
pH	6.95-7.05	6.97
Conductivity( $\mu\text{s}/\text{cm}$ )	N/A	N/A
Osmolality(mOsm/kg)	285-295	289

e.g. Meets the requirements set out in 'Guidelines for the Blood Transfusion Services in the United Kingdom' 7th Edition; as set out in chapter 12, 12.2.4, Miscellaneous reagents.

**Preservatives used:** Sodium azide (0.01% [w/v])

**Sterility:** Product is filtered through a 0.2 $\mu\text{m}$  filter into gamma irradiated bags

**Storage:** Room temperature

**Country of Origin:** United Kingdom

**Health and Safety:**

**Precautionary Notes**

Please download the Safety Data Sheet at <http://sourcebioscience.com>

**Signed for Quality Control:** 

**Date:** 6-APR-18

**Signed for Certificate Check:** 

**Date:** 6 APR 18