

CERTIFICATE OF ANALYSIS

Certificate of Analysis

Product name: Wash Buffer Base
Product code(s): 04-331
Batch Number(s): 74180402
Date Manufactured: 18 Apr 18
Expiry Date: 14 Dec 18
Appearance: Clear Liquid

Quality Control:

	Specification	This batch result
pH	7.80 - 8.20	7.82
Conductivity($\mu\text{s}/\text{cm}$)	N/A	N/A
Osmolality(mOsm/kg)	N/A	N/A

Manufacture Specification:
SPN319/1.1, effective 29/03/17

Preservatives used: None

Sterility: Terminally filtered with 0.2 μm filter in a grade C clean room and tested for sterility.

Storage: 2-25 °C

Country of Origin: United Kingdom

Health and Safety:

Precautionary Notes

Please download the Safety Data Sheet at <http://sourcebioscience.com>

Signed for Quality Control: 

Date: 09 May 18

Signed for Certificate Check: 

Date: 09 May 18