

## CERTIFICATE OF ANALYSIS

# Certificate of Analysis

**Product name:** Reagent Grade Water

**Product code(s):** 04-035

**Batch Number(s):** 18180588

**Date Manufactured:** 10 May 2018

**Expiry Date:** 09 May 2019

**Appearance:** Clear Liquid

**Quality Control:**

	Specification	This batch result
pH	6.95-7.05	N/A
Conductivity( $\mu\text{s}/\text{cm}$ )	<10	4.150
Osmolality(mOsm/kg)	285-295	N/A

**Preservatives used:** None

**Sterility:** No chemicals used in this preparation. Water is run through the reverse osmosis unit and collected without it being stored. It is collected straight into the appropriate bags or bottles with as little exposure to the atmosphere as possible

**Storage:** Room temperature

**Country of Origin:** United Kingdom

**Health and Safety:**

**Precautionary Notes**

Please download the Safety Data Sheet at <http://sourcebioscience.com>

**Signed for Quality Control:** 

**Date:** 10-MAY-18.

**Signed for Certificate Check:** 

**Date:** 10 MAY 18