

CERTIFICATE OF ANALYSIS

Certificate of Analysis

Product name: PBSS AF (Phosphate Buffered Serology Saline Azide-Free)

Product code(s): 04-103

Batch Number(s): 01171193

Date Manufactured: 22 Nov 17

Expiry Date: 21 Nov 19

Appearance: Clear Liquid

Quality Control:

	Specification	This batch result
pH	6.95-7.05	7.03
Conductivity($\mu\text{s}/\text{cm}$)	N/A	N/A
Osmolality(mOsm/kg)	285-295	288

Meets the requirements set out in 'Guidelines for the Blood Transfusion Services in the United Kingdom' 7th Edition; as set out in chapter 12, 12.2.4, Miscellaneous reagents.

Preservatives used: N/A

Sterility: Product is filtered through a 0.2 μm filter into gamma irradiated bags

Storage: Room temperature

Country of Origin: United Kingdom

Health and Safety:

Precautionary Notes

Please download the Safety Data Sheet at <http://sourcebioscience.com>

Signed for Quality Control:



Date: 23 Nov 17

Signed for Certificate Check:



Date: 23 Nov 17