

CERTIFICATE OF ANALYSIS

Certificate of Analysis

Product name: Heart Valve Cryopreservation Medium
Product code(s): 04-311
Batch Number(s): 71180408
Date Manufactured: 05 Apr 18
Expiry Date: 8 Apr 19
Appearance: Red translucent liquid

Quality Control:

	Specification	This batch result
pH	7.00-7.50	7.28
Conductivity($\mu\text{s}/\text{cm}$)	N/A	N/A
Osmolality(mOsm/kg)	N/A	N/A

Manufacturing specification:
SPN323/1, effective 28/03/17

Preservatives used: None

Sterility: Terminally filtered with 0.2 μm filter in grade B clean room and tested for sterility

Storage: Ambient

Country of Origin: United Kingdom

Health and Safety:

Precautionary Notes

Please download the Safety Data Sheet at <http://sourcebioscience.com>

Signed for Quality Control: 

Date: 17 MAY 18

Signed for Certificate Check: 

Date: 17 MAY 18