

CERTIFICATE OF ANALYSIS

Certificate of Analysis

Product name: LISS (Low Ionic Strength Saline)

Product code(s): 04-017

Batch Number(s): 03180623

Date Manufactured: 18 Jun 18

Expiry Date: 17 Jun 19

Appearance: Clear liquid

Quality Control:

	Specification	This batch result
pH	6.50 - 6.90	6.81
Conductivity($\mu\text{s}/\text{cm}$)	3500 - 3800	3534
Osmolality(mOsm/kg)	285 - 295	289

Preservatives used: Sodium azide (0.09% [w/v])

Sterility: N/A

Storage: Ambient

Country of Origin: United Kingdom

Health and Safety:

Precautionary Notes

Wear appropriate Personal Protective Equipment (PPE) when handling this reagent.

Please download the Safety Data Sheet at <http://sourcebioscience.com>

Signed for Quality Control:



Date: 21 Jun 18

Signed for Certificate Check:



Date: 21 Jun 18