

CERTIFICATE OF ANALYSIS

Certificate of Analysis

Product name: Frozen Cell Recovery Solution
Product code(s): 04-161BAG and 04-162BAG
Batch Number(s): 41180618
Date Manufactured: 12 Jun 18
Expiry Date: 28 Aug 19
Appearance: Clear Liquid

Quality Control:

	Specification	This batch result
pH	6.40 - 6.80	6.53
Conductivity($\mu\text{s}/\text{cm}$)	44,000 - 52,000	46,736
Osmolality(mOsm/kg)	1080 - 1120	1108

Preservatives used: EDTA (anti-fungal)

Sterility: Terminally filtered with 0.2 μm filter in a grade B clean room and tested for sterility.

Storage: 2 - 8°C

Country of Origin: United Kingdom

Health and Safety:

Precautionary Notes

Please download the Safety Data Sheet at <http://sourcebioscience.com>

Signed for Quality Control: 

Date: 29 Jun 18

Signed for Certificate Check: 

Date: 29 Jun 18