

## CERTIFICATE OF ANALYSIS

# Certificate of Analysis

**Product name:** LISS (Low Ionic Strength Saline)

**Product code(s):** 04-123

**Batch Number(s):** 03180725

**Date Manufactured:** 16 Jul 18

**Expiry Date:** 15 Jul 19

**Appearance:** Clear Liquid

**Quality Control:**

	Specification	This batch result
pH	6.50 - 6.90	6.85
Conductivity( $\mu\text{s}/\text{cm}$ )	3500 - 3800	3511
Osmolality(mOsm/kg)	285 - 295	291

**Preservatives used:** Sodium azide (0.09% [w/v])

**Sterility:** N/A

**Storage:** Ambient

**Country of Origin:** United Kingdom

**Health and Safety:**

**Precautionary Notes**

Wear appropriate Personal Protective Equipment (PPE) when handling this reagent.

Please download the Safety Data Sheet at <http://sourcebioscience.com>

**Signed for Quality Control:**



**Date:** 23 Jul 18

**Signed for Certificate Check:**



**Date:** 23 Jul 18