

CERTIFICATE OF ANALYSIS

Certificate of Analysis

Product name: Hypotonic Buffer Base

Product code(s): 04-333

Batch Number(s): 76180703

Date Manufactured: 08 Aug 18

Expiry Date: 07 Aug 19

Appearance: Clear Liquid

Quality Control:

	Specification	This batch result
pH	7.80 - 8.20	8.01
Conductivity($\mu\text{s}/\text{cm}$)	N/A	N/A
Osmolality(mOsm/kg)	N/A	N/A

Manufactured to Specification:
SPN320/1, dated 29 Mar 2017..

Preservatives used: None

Sterility: Terminally filtered with 0.2 μm filter in grade B clean room and tested for sterility

Storage: 2-25oC

Country of Origin: United Kingdom

Health and Safety:

Precautionary Notes


Please download the Safety Data Sheet at <http://sourcebioscience.com>

Signed for Quality Control:



Date: 22 Aug 18

Signed for Certificate Check:



Date: 22 Aug 18